INVIRO MEDICAL
SOLUTIONS FOR SAFETY®

1755 North Brown Road
Suite 150
Lawrenceville, GA 30043
Tel 678 405 4025
Fax 678 405 4044
www.inviromedical.com

Instructions for Use
InviroSNAP!® Safety Syringe: Fixed Needle

1. **Peel Package as indicated**, to allow proper removal of the InviroSNAP! Retractable Safety Syringe: Fixed Needle (InviroSNAP! Safety Syringe) and remove needle cap.

2. Prepare vial for medication draw up. **Do not push the syringe plunger past the first gradation mark**.

3. **Draw up medication** per facility protocol. (See Figure 1).

4. Prepare injection site and administer medication per facility protocol.

5. **To engage the safety mechanism, depress the plunger FULLY** to enable retraction of the needle. (See Figure 2).

6. **Pull back on plunger until it stops**: observe withdrawal of needle into barrel. (See Figure 3).

7. **Snap! off plunger** to permanently disable syringe. (See Figure 4).

8. **Discard immediately in sharps container** per facility protocol.

Figure 1

Figure 2

Figure 3

Figure 4

Caution

The InviroSNAP! Safety Syringe should not be used to take blood samples.

This device is sterile unless the package has been damaged or opened.

If unable to activate safety mechanism, discard syringe immediately into an approved sharps container.

This device is a single use item only.

SNAP!® syringe patents US: 5,263,933, 5,462,531, 5,520,649, 6,033,386, 6,117,113, 6,344,031 and additional foreign patents and patents pending. Additional applicable patents: US: 5263934, 5466226, 5380286, 5364359 and 6315760; Canada: 131718. Caution; US Law restricts this device to sale by or on the order of a physician. IFU_Fixed_Needle_nonStr Revised Oct 2007